

Certificate of Analysis

Anti-Human CD3 antibody

Cat No.: B104-0010 Lot No.: B1040010220310
Production Date: 10th March, 2022 Expiry Date: 9th March, 2025
Storage: -20 °C or below Page: 1 / 1

Quality control

	Specifications	Results
Appearance	Transparent to white solid	Pass
Sterility ^[1]	Negative	Pass
Purity ^[2]	≥ 95%	Pass
Concentration ^[3]	≥ 2.0 mg/mL	2.51 mg/mL
Endotoxin ^[4]	< 1.0 U/mg	Pass

<https://nobimpex.de/en/>

References

- [1] Direct Inoculation; Current Edition EP
- [2] SDS-PAGE
- [3] ELISA
- [4] Gel-clot technique; Current Edition EP

These signatures indicate the material mentioned above has met specifications and has been reviewed by the quality unit.

Reconstitution: Add 0.5ml ddH₂O to final concentration of 2mg/ml anti CD3 antibody. For further dilution, use PBS w/wo carrier protein, such as albumin.

Signature: 
Position: Approver
Issue date: 27th April, 2022